

Document reference: ATC-ML-TW-RP-19

**ICOS ATC Metrology Laboratory  
Evaluation report  
for the LICOR LI-7815 instrument**

Tests by L. Lienhardt  
Approved by O. Laurent  
Date: 2020-06-10

## Document history

Version	Date	Actions
1.0	2020-06-10	Creation

## Diffusion

- ATC internal
- ICOS community
- Public

## Repository

- Alfresco in Library/Documents/Common/ICOS-RI/ATC/MetrologyLab/Reports
- ICOS ATC website: <https://icos-atc.lsce.ipsl.fr/docs>

## Disclaimer

The contents of this document (including any attachments) may be privileged, confidential or copyrighted under applicable law and are intended solely for use by the intended recipient. The status is discussed only using the indicated version of the ICOS atmospheric station specifications.

## Contents

<b>1 Instrument references</b>	<b>4</b>
<b>2 Initialization Time</b>	<b>5</b>
<b>3 Warm restart</b>	<b>6</b>
<b>4 Cold restart</b>	<b>7</b>
<b>5 Continuous Measurement Repeatability (CMR) assessment</b>	<b>8</b>
<b>6 Short term stability and drift assessment</b>	<b>9</b>
<b>7 Short Term Repeatability (STR) assessment</b>	<b>10</b>
<b>8 Long Term Repeatability (LTR) assessment</b>	<b>11</b>
8.1 First period	11
8.2 Second period	11
8.3 Third period	12
8.4 Fourth period	12
8.5 Fifth period	13
8.6 Sixth period	13
8.7 Over all Long Term Repeatability	14
<b>9 Atmospheric pressure sensitivity</b>	<b>15</b>
<b>10 Inlet pressure sensitivity</b>	<b>16</b>
<b>11 Temperature sensitivity</b>	<b>18</b>
<b>12 Water vapor correction assessment</b>	<b>19</b>
12.1 Factory correction	19
12.2 Determination of H <sub>2</sub> O correction coefficients by the MLab	20
12.3 MLab correction	21
<b>13 Calibration</b>	<b>22</b>
<b>14 Linearity</b>	<b>23</b>
<b>15 Rise Time and Fall Time</b>	<b>24</b>
<b>16 Laboratory inter-comparison</b>	<b>25</b>
16.1 Without drying system	25
16.1.1 First period	25
16.1.2 Second period	27
16.1.3 Third period	29
16.1.4 Fourth period	31
16.2 Over all Long Term Repeatability	33
16.3 With drying system	35
16.3.1 First period	35
16.3.2 Second period	36
<b>17 Summary</b>	<b>37</b>

## 1 Instrument references

Owner	Reception Date	Departure date
LICOR	2019-09-23	2019-12-17


Brand	Model	S/N	Software release version
LICOR	LI-7815	TG15-01026	

ID	Associated documents	Reference	Date
AD1	Procedure of initial tests	ATC-ML-IT-PR-02-2.0	2016-10-14
AD2	ICOS atmospheric station specifications	ATC-GN-GN-SP-1.3	2017-11
AD3	Incoming control sheet	ATC-ML-IT-IC-04	
AD4	Follow-up sheet	ATC-ML-IT-FS-09	2019-12-02

In the following pages, we present the results of the tests performed at the ATC MLab. For more details about these tests, please refer to the procedure of initial tests [AD1]. For each test, we either show the results not corrected for the water vapor (w, e.g.  $\text{CO}_2w$ ) or corrected for the water vapor using the factory correction or the ATC correction (d, e.g.  $\text{CO}_2d$ ). Except for the temperature test, the laboratory temperature is regulated at  $22^\circ\text{C} \pm 2$ .

## 2 Initialization Time

**Methodology:** Measure continuously a tank filled with dry natural air after 1 hour of shutdown. No calibration applied. No rejected data. Target gas measured right after turning the instrument on.





	CO2 (min)
Initialization time	30

### 3 Warm restart

**Methodology:** Measure continuously a tank filled with dry natural air after several restart. No calibration applied. No rejected data. Target gas measured right after turning the instrument on.



1 minute off then 1 hour on (3 times) ; 1 hour off then 1 hour on (2 times)



	CO2 [ppb]
Maximum offset between two successive restarts	170

## 4 Cold restart



**Methodology:** Measure continuously a tank filled with dry natural air after 15 hours of shutdown. No calibration applied. No rejected data. Target gas measured right after turning the instrument on.



	CO2 [ppb]
Maximum offset between two successive restarts	240

## 5 Continuous Measurement Repeatability (CMR) assessment


**Methodology:** Measure continuously a tank filled with dry natural air during at least 25 hours. Look at data distribution for different integration times. First hour not taken into account (stabilization time). No calibration applied.



	CO2 [ppb]	H2O [%v]
Average of the standard deviations of raw data over a minute	43.19	0.00
Minute averaged data CMR Precision ( $1\sigma$ )	19.70	0.00
Hourly averaged data CMR Precision ( $1\sigma$ )	18.70	0.00
Minute averaged data CMR MaxDrift (peak to peak)	98.50	0.00
Hourly averaged data CMR MaxDrift (peak to peak)	66.00	0.00

## 6 Short term stability and drift assessment

**Methodology:** Measure continuously a tank filled with dry natural air during at least 25 hours. Calculate Allan deviations. First hour not taken into account (stabilization time). No calibration applied.




	CO2
Optimum Averaging time (s)*	1800
Optimum Allan deviation [ppb]*	3
Allan deviation at 1 min [ppb]	5
Allan deviation at 5 min [ppb]	4
Allan deviation at 10 min [ppb]	4
Allan deviation at 15 min [ppb]	4
Allan Deviation at 1 hr [ppb]	4

\* The optimum is searched in the first one hour window.

## 7 Short Term Repeatability (STR) assessment

**Methodology:** Measure a tank filled with dry natural air for 15 min and wet ambient air for 5 minutes alternatively 10 times. For each period of tank measurement, calculate a mean value (discard the first minutes for stabilization). Look at the dispersion ( $1\sigma$ ) of the mean values (10 points). No calibration applied.




	CO2 [ppb]
Short term repeatability ( $1\sigma$ , 9 minute average raw data)	4.00
MaxDrift (peak to peak, 9 minute average raw data)	14.00

## 8 Long Term Repeatability (LTR) assessment


**Methodology:** Measure alternatively over 72 hours a tank filled with dry natural air for 30 minutes and 270 minutes of wet ambient air. For each period of tank measurement, calculate a mean value (discard the first minutes for stabilization). Look at the dispersion ( $1\sigma$ ) of the mean values. No calibration applied.

### 8.1 First period




	CO2 [ppb]
Long term repeatability ( $1\sigma$ , 10 minute average raw data)	35.00
MaxDrift (peak to peak, 10 minute average raw data)	106.00

### 8.2 Second period




	CO2 [ppb]
Long term repeatability ( $1\sigma$ , 10 minute average raw data)	45.00
MaxDrift (peak to peak, 10 minute average raw data)	123.00

### 8.3 Third period




	CO2 [ppb]
Long term repeatability ( $1\sigma$ , 10 minute average raw data)	25.00
MaxDrift (peak to peak, 10 minute average raw data)	84.00

### 8.4 Fourth period




	CO2 [ppb]
Long term repeatability ( $1\sigma$ , 10 minute average raw data)	32.00
MaxDrift (peak to peak, 10 minute average raw data)	117.00

## 8.5 Fifth period



	CO2 [ppb]
Long term repeatability ( $1\sigma$ , 10 minute average raw data)	23.00
MaxDrift (peak to peak, 10 minute average raw data)	78.00


## 8.6 Sixth period



	CO2 [ppb]
Long term repeatability ( $1\sigma$ , 10 minute average raw data)	26.00
MaxDrift (peak to peak, 10 minute average raw data)	73.00

## 8.7 Over all Long Term Repeatability


Restarts performed from November 27th to 28th



	CO2 [ppb]
Long term repeatability ( $1\sigma$ , 10 minute average raw data)	756.00
MaxDrift (peak to peak, 10 minute average raw data)	2088.00

## 9 Atmospheric pressure sensitivity

**Methodology:** Measure alternatively over 72 hours a tank filled with dry natural air for 30 minutes and 270 minutes of wet ambient air. For each period of tank measurement, calculate a mean value (last 10 minutes) and look at the correlation of the tank measurement with atmospheric pressure (AP) variation.




	CO2
Atmospheric pressure sensitivity (ppb/hPa)	NS

Not significant (NS) if  $R^2 < 0.5$  or the absolute value of the slope lower than 1 and 0.02 for  $\text{CO}_2$  and  $\text{CH}_4$  respectively

## 10 Inlet pressure sensitivity


**Methodology:** Measure continuously a tank filled with dry natural air through an electronic pressure controller at the instrument inlet. Change sequentially (step of 20 minutes) the instrument inlet pressure (maximum range from 1200 mbar absolute to 600 mbar absolute) thanks to the pressure controller. The valid range is evaluated as the range where CO<sub>2</sub> mixing ratios are  $pm$  0.02ppm from the mixing ratio at atmospheric pressure..





## 11 Temperature sensitivity

**Methodology:** Measure a tank filled with dry natural air while changing the ambient temperature ( $T_{amb}$ ). Look at the correlation of the measurement stability with the instrument internal temperature ( $T_{cav}$ ).



	CO2
Ambient temperature sensitivity (ppb/°C)	NS


Not significant (NS) if  $R^2 < 0.45$  or the absolute value of the slope lower than 5 and 0.1 ppb/°C for  $CO_2$  and  $CH_4$  respectively

## 12 Water vapor correction assessment

**Methodology:** Measure a tank filled with dry natural air during at least 1h. Then humidify by 20 minute steps the tank gas at 0.25/0.5/0.75/1/1.25/1.5/1.75/2/2.5/3 %v of water vapor. Finally, stop humidifying and measure the tank filled with dry natural air during more than 1 hour. Repeat the experiment at least twice, usually three times. Check the water vapor correction bias depending on the H<sub>2</sub>O level. Determine an optimized water vapor correction bias.

$$H_2O \text{ correction bias} = C_{\text{humidified gas, water vapor corrected}} - C_{\text{not humidified gas}} \quad (1)$$

### 12.1 Factory correction




## 12.2 Determination of H<sub>2</sub>O correction coefficients by the MLab

$$C_r = \frac{C_{wet}}{C_{dry}} = 1 + I_1 * H_2O_r + I_2 * H_2O_r^2 \quad (2)$$

with H<sub>2</sub>O<sub>r</sub>: Instrument reported value (not calibrated). The “calibrated” H<sub>2</sub>O value is:


$$H_2O = 0.772 * H_2O_r + 0.019493 * H_2O_r^2 \quad (3)$$



H<sub>2</sub>O correction coefficients determined by ATC

	I1	I2
1	-5.680e-04	3.223e-04


### 12.3 MLab correction



### 13 Calibration

**Methodology:** Measure 4 times (during 20 minutes each time) 4 standards filled with known CO<sub>2</sub>, CH<sub>4</sub> and CO concentrations. Compare reported values from the instrument and assigned values. Determine calibration functions. The residuals shown are the residuals from the calibration fit ( $C_{Assigned} - C_{Reported} = f(C_{Assigned})$ ). Check the instrument drift. The value in the table is evaluated by calculating the temporal regression of the average differences (Assigned-Reported) for each calibration episode.


CO2



	CO2
Calibration drift trend (ppb/month)	-799.9
Maximum residual from linear fit on calibration range [ppb]	24

## 14 Linearity

**Methodology:** Measure 4 times (during 20 minutes each time) 6 standards filled with known CO<sub>2</sub>, CH<sub>4</sub> and CO concentrations within the range guaranteed by the manufacturer. The first minutes are not taken into account (stabilization time). No calibration applied.




	CO2
Maximum residual from linear fit on extended mole fraction range [ppb]	19
Maximum residual from linear fit on extended mole fraction range [%]	0.005


## 15 Rise Time and Fall Time

**Methodology:** Measure the rise time and fall time between two different tanks during a calibration.

Rise time : Time taken for the response to rise from 10% to 90% of its final normalized value



Fall time : Time taken for the response to fall from 90% to 10% of its final normalized value



	CO2 (sec)
Rise time	<2
Fall time	<2


## 16 Laboratory inter-comparison

### 16.1 Without drying system

**Methodology:** Measure ambient air without drying system. Compare with a reference instrument with drying system. The 2 instruments are calibrated against the same set of calibration tanks. They are each equipped with a dedicated sampling line. If the MLab reference instrument is unavailable then the reference instrument is the instrument tested in parallel. In this case, they use the same sampling line and the ATC water vapor correction is applied to the reference. A target gas is measured on both instruments for quality control.


#### 16.1.1 First period

##### 16.1.1.1 Factory water vapor correction



	CO2 [ppb]
Observed bias in ambient air:	
mean difference (LI-7815 - Ref Instrument)	-154
H <sub>2</sub> O correction bias estimated by ATC for the mean H <sub>2</sub> O during the test ( 11883 ppm H <sub>2</sub> O)	-94
Remaining bias (not related to H <sub>2</sub> O correction)	-60


16.1.1.2 Water vapor correction coefficients determined by ATC



	CO2 [ppb]
Observed bias : mean difference (LI-7815 - Reference)	-53
H <sub>2</sub> O correction bias estimated by ATC for the mean H <sub>2</sub> O during the test ( 11883 ppm H <sub>2</sub> O)	-9
Remaining bias (not related to H <sub>2</sub> O correction)	-44


## 16.1.2 Second period

### 16.1.2.1 Factory water vapor correction



	CO2 [ppb]
Observed bias in ambient air:	
mean difference (LI-7815 - Ref Instrument)	-90
H <sub>2</sub> O correction bias estimated by ATC	
for the mean H <sub>2</sub> O during the test ( 12149 ppm H <sub>2</sub> O)	-94
Remaining bias (not related to H <sub>2</sub> O correction)	4


16.1.2.2 Water vapor correction coefficients determined by ATC



	CO2 [ppb]
Observed bias : mean difference (LI-7815 - Reference)	5
H <sub>2</sub> O correction bias estimated by ATC for the mean H <sub>2</sub> O during the test ( 12149 ppm H <sub>2</sub> O)	-9
Remaining bias (not related to H <sub>2</sub> O correction)	14


### 16.1.3 Third period

#### 16.1.3.1 Factory water vapor correction



	CO2 [ppb]
Observed bias in ambient air:	
mean difference (LI-7815 - Ref Instrument)	-47
H <sub>2</sub> O correction bias estimated by ATC for the mean H <sub>2</sub> O during the test ( 4395 ppm H <sub>2</sub> O)	-101
Remaining bias (not related to H <sub>2</sub> O correction)	54


16.1.3.2 Water vapor correction coefficients determined by ATC



	CO2 [ppb]
Observed bias : mean difference (LI-7815 - Reference)	42
H <sub>2</sub> O correction bias estimated by ATC for the mean H <sub>2</sub> O during the test ( 4395 ppm H <sub>2</sub> O)	-21
Remaining bias (not related to H <sub>2</sub> O correction)	63


## 16.1.4 Fourth period

### 16.1.4.1 Factory water vapor correction



	CO2 [ppb]
Observed bias in ambient air:	
mean difference (LI-7815 - Ref Instrument)	-188
H <sub>2</sub> O correction bias estimated by ATC for the mean H <sub>2</sub> O during the test ( 7721 ppm H <sub>2</sub> O)	-117
Remaining bias (not related to H <sub>2</sub> O correction)	-71
Observed bias on TGT (dry air):	
mean difference (LI-7815 - Ref Instrument)	-17


16.1.4.2 Water vapor correction coefficients determined by ATC



	CO2 [ppb]
Observed bias : mean difference (LI-7815 - Reference)	-77
H <sub>2</sub> O correction bias estimated by ATC for the mean H <sub>2</sub> O during the test ( 7721 ppm H <sub>2</sub> O)	-19
Remaining bias (not related to H <sub>2</sub> O correction)	-58
Observed bias on TGT (dry air): mean difference (LI-7815 - Ref Instrument)	-17


## 16.2 Over all Long Term Repeatability

### 16.2.0.1 Factory water vapor correction



	CO2 [ppb]
Observed bias in ambient air:	
mean difference (LI-7815 - Ref Instrument)	-140
H <sub>2</sub> O correction bias estimated by ATC for the mean H <sub>2</sub> O during the test ( 11155 ppm H <sub>2</sub> O)	-94
Remaining bias (not related to H <sub>2</sub> O correction)	-46

### 16.2.0.2 Water vapor correction coefficients determined by ATC




	CO2 [ppb]
Observed bias : mean difference (LI-7815 - Reference)	-40
H <sub>2</sub> O correction bias estimated by ATC for the mean H <sub>2</sub> O during the test ( 11155 ppm H <sub>2</sub> O)	-9
Remaining bias (not related to H <sub>2</sub> O correction)	-31

### 16.3 With drying system


**Methodology:** Measure ambient air with drying system. Compare with a reference instrument. The 2 instruments are calibrated against the same set of calibration tanks. They are equipped with a dedicated sampling line. In this case, they use the same sampling line. A target gas is measured on both instruments for quality control.

#### 16.3.1 First period



	CO2 [ppb]
Observed bias : mean difference (LI-7815 - Reference)	127
Intrinsic bias estimated in wet air conditions (ATC correction)	-31
Observed bias on TGT (dry air): mean difference (LI-7815 - Ref Instrument)	20

16.3.2 Second period



	CO2 [ppb]
Observed bias : mean difference (LI-7815 - Reference)	91
Intrinsic bias estimated in wet air conditions (ATC correction)	-31

## 17 Summary

For legibility purposes, the results are split into tables by species.  
Only status in bold are taken into account for the final status.

	CO <sub>2</sub>			
	Spec	ATC	unit	Status
Field CMR (average on min sd)	-	43.1893163774142	ppb	-
<b>Minute CMR (1<math>\sigma</math>)</b>	<50	<b>19.7</b>	ppb	<b>Pass</b>
<b>Hourly CMR (1<math>\sigma</math>)</b>	<25	<b>18.7</b>	ppb	<b>Pass</b>
Minute CMR MaxDrift (peak to peak)	<200	98.5	ppb	Pass
Hourly CMR MaxDrift (peak to peak)	<150	66	ppb	Pass
<b>LTR-1 (1<math>\sigma</math>, 10 min avr raw data)</b>	<50	<b>35</b>	ppb	<b>Pass</b>
LTR-1 MaxDrift (peak to peak)	<200	106	ppb	Pass
<b>LTR-2 (1<math>\sigma</math>, 10 min avr raw data)</b>	<50	45	ppb	<b>Pass</b>
LTR-2 MaxDrift (peak to peak)	<200	123	ppb	Pass
<b>LTR-3 (1<math>\sigma</math>, 10 min avr raw data)</b>	<50	25	ppb	<b>Pass</b>
LTR-3 MaxDrift (peak to peak)	<200	84	ppb	Pass
<b>LTR-4 (1<math>\sigma</math>, 10 min avr raw data)</b>	<50	32	ppb	<b>Pass</b>
LTR-4 MaxDrift (peak to peak)	<200	117	ppb	Pass
STR (1 $\sigma$ , 9 min avr raw data)	-	4	ppb	-
Atm. pressure sensitivity	-	NS	ppb/hPa	-
Temperature sensitivity	-	NS	ppb/°C	-
Max res from fit in cal range	-	24	ppb	-
Max res from fit in extended range	-	19	ppb	-
Max res from fit in extended range 2	-	0.005	%	-
Calibration drift trend	-	-799.9	ppb/month	-
Water vapor corr: max bias ATC	-	31.28	ppb	-
Water vapor corr: max bias Factory	-	436.39	ppb	-
Water vapor correction I1	-	-5.680e-04	-	-
Water vapor correction I2	-	3.223e-04	-	-